



CODE **GT0109**
 OVERALL DIMENSIONS 560x620x1000 mm
 POWER 0.25 kW
 WEIGHT 80 Kg

DEEP UNGLAZED ABRASIMETER

STANDARDS

UNI EN ISO 10545-6, UNI EN 14617- 4, UNI EN 1344, UNE 127024:1999 EX

DESCRIPTION

Laboratory instrument to determine the resistance to deep abrasion in the unglazed ceramic tiles obtained by means a steel disc Ø 200 x 10 mm and abrasive material - as corundum grain size F80 - that falls between the disk and the specimen.

*The instruments are calibrated according to the UNI EN ISO 10545-6.
 Other calibration on demand.*

TECHNICAL SPECIFICATIONS

- Steel structure powder epoxy painted
- Plexiglas safety protection
- Control flow abrasive outlet: Is done to according to the standard
- Steel disc Ø 200 x 10 mm
- Hardness disc is done according to the standard
- Supply: 230 V - 50/60 Hz single phase (other voltage on request)
- Electronic counter
- Maximum size of the sample: 180x180x80 mm
- Ready for connection to a vacuum cleaner

EQUIPMENT

- Calibration plate by transparent fused silica 100x10 x3 mm
- 5 Kg Corundum (aluminium fused oxide) white, grain size F80

ACCESSORIES AND SPARE PARTS

GT1367	Pack of 25 kg Corundum (aluminium fused oxide) white, grain size F80
GT0110	Calibration plate by transparent fused silica 100x100x3mm
GT0111	Steel disc Ø 200 x 10 mm
GT0121	Conversion kit from DEEP/UNGLAZED to DEEP/STONE
GT1316	Kit for variation in number of revolutions
GT0343	Digital caliper DC-150 (length 150 mm)
GT1722	Kit UNI EN 14617-4